



Certificate of Conformance
SAMPLE ONLY



CCG PART #: 00119146
CUSTOMER: SAMPLE
CUSTOMER PO #: XXXXXXXX
CUSTOMER ITEM #:
QUANTITY: 1 **UOM:** CS
LOT #: XXXXXXXX
DOM: XX/XX/XXXX **EXPIRATION:** 3 Years from DOM
DESCRIPTION: F-PGFP-001L | FlowTainer 1L PETG Erlenmeyer Flask, Flat Cap, 12/CS

MATERIALS OF CONSTRUCTION:

Manufactured from PETG

BIOCOMPATIBILITY:

This product is certified to meet USP 85., Meets USP 87 Biological Reactivity Tests, In Vitro., Meets USP 88 Biological Reactivity Tests, Class VI, In Vivo., Resin is non-hemolytic in accordance with ISO 10993-4., Resin is non-cytotoxic in accordance with ISO 10993-5., Resin is in accordance with ISO 10993-6., Resin is in accordance with ISO 10993-10., Resin is in accordance with ISO 10993-11.

TSE/BSE/ADCF STATEMENT:

No animal products or by-products are used/added in mfg.

REACH/RoHS:

Material is compliant with REACH/RoHS requirements.

GAMMA-IRRADIATION COMPATIBILITY:

May be exposed to Gamma-Irradiation to a total of 45kGy.

USP<85> BACTERIAL ENDOTOXIN:

This product is certified to meet USP 85.

USP 788 :

This product is certified to meet USP 788.

PHYSIOCHEMICAL COMPATIBILITY:

Material is certified to meet USP <661>.

MANUFACTURING ENVIRONMENT:

Manufactured in ISO 7 Cleanroom per ISO 14644.

Supplied By: Carolina Components Group, 1001 Hill Dr. Suite 110, Durham, NC 27703

Quality Assurance,

Carolina Components Group

Date: 11/15/2023