



**Certificate of Conformance
SAMPLE ONLY**



CCG PART #: 00118819
CUSTOMER: SAMPLE
CUSTOMER PO #: XXXXXXXX
CUSTOMER ITEM #:
QUANTITY: 5 **UOM:** PK
LOT #: XXXXXXXX
DOM: XX/XX/XXXX **EXPIRATION:** 5 Years from DOM
DESCRIPTION: F-SN-2.0 | FlowLinX Sampling Needle, 2.0mm

MATERIALS OF CONSTRUCTION:

Manufactured from Platinum Silicone LSR., Manufactured from Formosa 5090R polypropylene resin.

BIOCOMPATIBILITY:

Meets USP 88 Biological Reactivity Tests, Class VI, In Vivo., Resin is non-hemolytic in accordance with ISO 10993-4., Resin is non-cytotoxic in accordance with ISO 10993-5.

TSE/BSE/ADCF STATEMENT:

No animal products or by-products are used/added in mfg.

REACH/RoHS:

Material is compliant with REACH/RoHS requirements.

GAMMA-IRRADIATION COMPATIBILITY:

May be exposed to Gamma-Irradiation to a total of 45kGy.

FOOD CONTACT APPLICATIONS:

Compliant with FDA 21 CFR 177.2600.

USP<85> BACTERIAL ENDOTOXIN:

This product is certified to meet USP 85.

MANUFACTURING ENVIRONMENT:

Manufactured in ISO 8 Cleanroom per ISO 14644.

Supplied By: Carolina Components Group, 1001 Hill Dr. Suite 110, Durham, NC 27703

Quality Assurance,

Carolina Components Group

Date: 10/12/2023