



Certificate of Conformance  
SAMPLE ONLY



**CCG PART #:** 00109871  
**CUSTOMER:** SAMPLE  
**CUSTOMER PO #:** XXXXXXXX  
**CUSTOMER ITEM #:**  
**QUANTITY:** 25 **UOM:** PK  
**LOT #:** XXXXXXXX  
**DOM:** XX/XX/XXXX **EXPIRATION:** 5 Years from DOM  
**DESCRIPTION:** F-TC-0500M | FlowLinX 3/4" (Mini) TC x 1/2" Barb

**MATERIALS OF CONSTRUCTION:**

Manufactured from P5M6K-080 polypropylene resin.

**BIOCOMPATIBILITY:**

Meets USP 88 Biological Reactivity Tests, Class VI, In Vivo., Resin is non-cytotoxic in accordance with ISO 10993-5.

**TSE/BSE/ADCF STATEMENT:**

No animal products or by-products are used/added in mfg.

**REACH/RoHS:**

Material is compliant with REACH/RoHS requirements.

**GAMMA-IRRADIATION COMPATIBILITY:**

May be exposed to Gamma-Irradiation to a total of 50kGy.

**FOOD CONTACT APPLICATIONS:**

Compliant with FDA 21 CFR 177.1520 ( c ) 3.1a.

**USP<85> BACTERIAL ENDOTOXIN:**

This product is certified to meet USP 85.

**MANUFACTURING ENVIRONMENT:**

Manufactured in ISO 8 Cleanroom per ISO 14644.

**Supplied By:** Carolina Components Group, 1001 Hill Dr. Suite 110, Durham, NC 27703

**Quality Assurance,**

**Carolina Components Group**

*Date: 10/04/2023*